

- 1. SUBSTRATE UV GRADE FUSED SILICA
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 ARCSEC
- 3. COATING (APPLY ACROSS COATING APERTURE)


PERFORMANCE SPECIFICATIONS @ 0° AOI T(AVG): >85% FROM 545 - 2000nm T(AVG): < 1% FROM 200 - 505nm T(ABS): = 50% 525nm @ 45° AOI

S1: MULTILAYER DIELECTRIC COATING S2: SINGLE LAYER MgF2

## 4. FINE GRIND SURFACE

- 5. TRANSMITTED WAVEFRONT DISTORTION RMS@ 632.8nm ≤ ¼ WAVE
- 6. ROHS COMPLIANT





## LONGPASS FILTER

Т

## SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

\_

	S1	\$2	]			Edmund Optic	<b>C</b> ®
SHAPE	PLANO	PLANO					5
SURFACE QUALITY	40-20	40-20		1		OD2, Ø50mm, 525nm LONGPASS FIL	TER
CLEAR APERTURE	80%	80%	THIRD ANGLE - PROJECTION	$\bigcirc \bigcirc$	TITLE		
COATING APERTURE	Ø40.00	Ø40.00	-				
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	64634	SHEET 1 OF 1