NOTES:

1. SUBSTRATE FUSED SILICA

- 2. SURFACE S2 TO BE PARALLEL TO SURFACE S1 TO WITHIN <3 ARCMIN
- 3. COATING (APPLY ACROSS COATING APERTURE):

S1: 266 HR Coating R (ABS) > 99.5% @ 266nm @ 45° AOI R (AVG) > 99.5% @ 263 - 268nm @ 45° AOI

DAMAGE THRESHOLD, PULSED: 2.5 J/cm², 20ns, 20Hz @ 266nm CW: 1MW/cm² @ 266nm


S2: NONE


4. FINE GRIND SURFACE

 POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE

6. CLEAR APERTURE AND COATING APERTURE ARE CENTERED ON SURFACES

APPLY ARROW ON EDGE WITH LASER ETCH, PENCIL, OR PERMANENT INK POINTING TOWARDS SURFACE \$1





FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	S2						
SHAPE	PLANO	PLANO					nd Optic	`∩®
SURFACE QUALITY	10-5	COMMERCIAL POLISH					na Optic	,5°
SURFACE FLATNESS	0.10 WAVE	N/A		1		Ø50.8mm x 9.53mm	266nm NdYAG MI	
MIN CLEAR APERTURE	Ø40.64	N/A	THIRD ANGLE PROJECTION		TITLE	Ø50.8mm x 9.53mm,266nm, NdYAG MIRROR 45° AOI		
MIN COATING APERTURE	Ø45.72	N/A						CHEET
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	38831		SHEET 1 OF