

## NOTES:

- 1. SUBSTRATE: VACUUM UV GRADE CaF2
- 2. SURFACE S2 TO BE PARALLEL WITH S1 TO <1 arcmin
- 3. COATING (APPLY ACROSS CLEAR APERTURE) \$1 & \$2: R(AVG) < 1% FROM 1650-3000nm @ 0° AOI R(ABS) < 2% FROM 1650-3000nm @ 0° AOI





SHEET 1 OF 1



## FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	40-20	40-20
CLEAR APERTURE	Ø18.00	Ø18.00
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

				<b>Eamuna Optic</b>	S
THIRD ANGLE PROJECTION		TITLE	Ø20mm, BBAR(1650 - 3000nm), CALCIUM FLUORIDE WINDOW		
	ALL DIMS IN	mm	DWG NO	15966	SHEE 1 OF