NOTES:

1. SUBSTRATE MATERIAL: FLOAT GLASS


2. COATING (APPLOW ACROSS CLEAR APERTURE): \$1: ALUMINUM


S2: NONE


3. BLAZE ANGLE: 8° 36' NOMINAL BLAZE WAVELENGTH: 250nm GROOVE DENSITY: 1200 GROOVES/mm

4.\ EDGES GROUND

5. Rohs Compliant







FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	Edmund	Optics®
L.	1 1 0 1 TM 10	200 0

				1.		Richardson Granings 1200 Grooves, 23 x	
	\$1	\$2	THIRD ANGLE PROJECTION	THIRD ANGLE PROJECTION		25mm, 250nm, Plane Ruled Reflection	
SHAPE	PLANO	PLANO				G	rating
CLEAR APERTURE	22.5 x 22.5	22.5 x 22.5	ALL DIMS IN	mm	DWG NO	15761	SHEET 1 OF 1